



## **Polymer Chip for Amperometric Detection of Neurotransmitter Release from Single Cells**

**Larsen, Simon Tyllsgaard; Tanzi, Simone; Taboryski, Rafael J.**

*Publication date:*  
2011

[Link back to DTU Orbit](#)

*Citation (APA):*

Larsen, S. T., Tanzi, S., & Taboryski, R. J. (2011). *Polymer Chip for Amperometric Detection of Neurotransmitter Release from Single Cells*. Poster session presented at PITTCON 2011, Atlanta, Georgia, USA.

---

### **General rights**

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.



# Polymer Chip for Amperometric Detection of Neurotransmitter Release from Single Cells

Simon Tylsgaard Larsen, Simone Tanzi and Rafael Taboryski

## Objective

Secretion of signaling molecules by cellular exocytosis is a key mechanism in many communication processes between living cells. Single exocytotic events can be studied by electrochemical sensing techniques using microelectrodes. The goal of this project is to develop a polymer microchip where neuroactive cells can be easily trapped near microelectrodes and to investigate the potential of the conductive polymer poly(3,4-ethylenedioxythiophene) (PEDOT) to work as microelectrode material for electrochemical detection of the neurotransmitter dopamine.

## Chip design and cell clamping

Microchips were made by combining two injection molded TOPAS parts. The upper part contained microfluidic channels and openings while the lower part was covered with PEDOT




Figure 1: a) Photomask designs for microchip fabrication. Microfluidic channels are shown in green (dark green: 50  $\mu\text{m}$  deep channels, light green: 5  $\mu\text{m}$  deep channels). Electrodes are shown in red. The large circles are chip openings. b) Illustration of cell clamping at a "mouse hole". The blue arrow indicates suction. c) Living PC12 cells can be trapped at the channel junctions. This microscope image shows a trapped polystyrene microbead of approximately same size as PC12 cells ( $r = 7.5 \mu\text{m}$ ). d) Microscope image of two PEDOT electrodes aligned with a channel junction.

A polymer microchip for recording of neurotransmitter release from living neuroactive PC 12 cells was developed. Electrochemical studies show that exocytotic dopamine signals can be recorded using conducting polymer poly(3,4-ethylenedioxythiophene) (PEDOT) microelectrodes. Microchips were made by combining two injection molded TOPAS parts. Nickel mold inserts for the injection molding process were electroformed from Si master templates containing the microfluidic channel structures.

microelectrodes. Living neuroactive PC 12 cells were successfully trapped at a specified site inside a chip using a so-called mouse hole configuration.

## Pedot microelectrodes

The conductive polymer PEDOT was used as electrode material. Cyclic voltammetry experiments showed that the neurotransmitter dopamine can be electrochemically detected on PEDOT electrodes (fig 2). By making the electrode surface sufficiently small noise levels below 2 pA can be




Figure 2: a) Current trace recorded at a PEDOT microelectrode. Spikes result from oxidation of dopamine being ejected into the phosphate buffer through a micropipette ( $d=2 \mu\text{m}$ ) situated close to the electrode. b) Cyclic voltammogram showing the oxidation and reduction of dopamine on a PEDOT electrode. Dopamine was diluted in PBS buffer. Concentration: 1 mM, scan rate: 20 mV/s, electrode dimensions: 4  $\mu\text{m}$  X 700  $\mu\text{m}$ .




Figure 3: a) Noise versus electrode area. The noise was calculated as the root mean square noise of a current trace with the electrode being exposed to phosphate buffer. b) Microscope image of a 2 micron wide PEDOT electrode.

reached. This is an acceptable noise value for measuring current peaks from single exocytotic events.

## Perspectives

One of the limitations of current technology used for measuring exocytotic signals electrochemically is the low through-put. We hope with this polymer chip platform to develop a device that is both cheap and capable of automatically measuring exocytotic activity of several cells in one sample.

